

PAPER

Automated detection of breast cancer in resected specimens with fluorescence lifetime imaging

To cite this article: Jennifer E Phipps *et al* 2018 *Phys. Med. Biol.* **63** 015003

View the [article online](#) for updates and enhancements.

Related content

- [Calibration and analysis of a multimodal micro-CT and structured light imaging system for the evaluation of excised breast tissue](#)
David M McClatchy III, Elizabeth J Rizzo, Jeff Meganck *et al.*
- [Towards intra-operative diagnosis of tumours during breast conserving surgery by selective-sampling Raman micro-spectroscopy](#)
Kenny Kong, Fazliyana Zaabar, Emad Rakha *et al.*
- [A review of Raman spectroscopy advances with an emphasis on clinical translation challenges in oncology](#)
Michael Jermyn, Joannie Desroches, Kelly Aubertin *et al.*




Time-saving, Comprehensive QA for **Multi-Metastasis SRS**

Benefits:

- Receives up to 29 Radiochromic films
- Enables precise and convenient film placement and registration
- Precise laser cut film packs available separately
- Proprietary material assures linear attenuation within 1% of brain tissue
- Compatible with ExacTrac® system

*ExacTrac® is a registered trademark of Brainlab (Munich, DE)





PAPER

Automated detection of breast cancer in resected specimens with fluorescence lifetime imaging

Jennifer E Phipps¹, Dimitris Gorpas^{1,5}, Jakob Unger¹, Morgan Darrow², Richard J Bold³ and Laura Marcu^{1,4}¹ Biomedical Engineering Department, University of California, Davis, CA 95616, United States of America² Department of Pathology and Laboratory Medicine, University of California Davis Health System, Sacramento, CA 95817, United States of America³ Department of Surgery, University of California Davis Health System, Sacramento, CA 95817, United States of America⁴ Author to whom any correspondence should be addressed.E-mail: lmarcu@ucdavis.edu**Keywords:** fluorescence lifetime imaging, breast cancer, support vector machinesSupplementary material for this article is available [online](#)RECEIVED
22 August 2017REVISED
28 October 2017ACCEPTED FOR PUBLICATION
3 November 2017PUBLISHED
14 December 2017**Abstract**

Re-excision rates for breast cancer lumpectomy procedures are currently nearly 25% due to surgeons relying on inaccurate or incomplete methods of evaluating specimen margins. The objective of this study was to determine if cancer could be automatically detected in breast specimens from mastectomy and lumpectomy procedures by a classification algorithm that incorporated parameters derived from fluorescence lifetime imaging (FLIm). This study generated a database of co-registered histologic sections and FLIm data from breast cancer specimens ($N = 20$) and a support vector machine (SVM) classification algorithm able to automatically detect cancerous, fibrous, and adipose breast tissue. Classification accuracies were greater than 97% for automated detection of cancerous, fibrous, and adipose tissue from breast cancer specimens. The classification worked equally well for specimens scanned by hand or with a mechanical stage, demonstrating that the system could be used during surgery or on excised specimens. The ability of this technique to simply discriminate between cancerous and normal breast tissue, in particular to distinguish fibrous breast tissue from tumor, which is notoriously challenging for optical techniques, leads to the conclusion that FLIm has great potential to assess breast cancer margins. Identification of positive margins before waiting for complete histologic analysis could significantly reduce breast cancer re-excision rates.

Introduction

Recent statistics indicate that breast cancer is the leading cause of cancer-related death and the 2nd most diagnosed cancer for women in the United States and is the most common cancer in women worldwide (Fitzmaurice *et al* 2016). Currently, an American woman has a 1 in 9 chance of developing breast cancer during her lifetime (Fitzmaurice *et al* 2016). Breast-conserving surgery (lumpectomy) followed by radiation is the standard-of-care surgical intervention for early-stage cancer and is as effective as mastectomy in many cases (O'Kelly Priddy *et al* 2015). From a 2009 study, 37.9% of 1459 lumpectomy procedures resulted in positive margins (Morrow *et al* 2009) and 50% of reoperations due to positive margin findings did not find residual tumor (Azu *et al* 2010). Additionally, positive margins are correlated with a significant increase in ipsilateral breast tumor regional recurrence (Houssami *et al* 2014) and reoperation is associated with greater physical and emotional trauma to the patient, a higher incidence of complications, and poorer cosmetic outcomes (St John *et al* 2017). The most accurate methods to assess tumor margins are cytology and frozen sections, both of which require significant time and cost and thus are not commonly performed (St John *et al* 2017). A fast, cost-effective and accurate way to assess breast cancer margins intraoperatively or immediately following resection is in high demand.

Optical techniques provide a means to non-destructively probe tissue composition, making them safe for intraoperative use. Studies have been done to determine the capability of several optical techniques to poten-

⁵ Current affiliation: Helmholtz Zentrum München, Institute of Biological and Medical Imaging, Munich, Germany

tially diagnose breast cancer specimens. This includes diffuse reflectance spectroscopy (DRS) (Brown *et al* 2010, Keller *et al* 2010, de Boer *et al* 2016), diffuse optical spectroscopy (DOS) (Nichols *et al* 2017), Raman spectroscopy (Kong *et al* 2014), fluorescence spectroscopy (Keller *et al* 2010), optical coherence tomography (OCT) (Nguyen *et al* 2009, Erickson-Bhatt *et al* 2015, Zysk *et al* 2015), optical coherence micro-elastography (Allen *et al* 2016), autofluorescence lifetime microscopy (Sharma *et al* 2012), and photoacoustic microscopy (Wong *et al* 2017). Moreover, few of these have been used to assess margins intraoperatively (i.e. DRS, DOS, and OCT), but none have been widely adopted into regular clinical practice due to inherent limitations. For example, while fibrous tissue will appear more uniform with OCT than tumor (Erickson-Bhatt *et al* 2015), OCT still has limited ability to distinguish between cancerous and fibrous breast tissue due to potentially similar structural features of these tissue types (Nguyen *et al* 2009). Also, while a careful study of normal and cancerous breast tissue has been performed with DRS (Kennedy *et al* 2016), an automated method to distinguish cancer based on DRS measurements has not been demonstrated. Furthermore, a recent meta-analysis of intraoperative margin assessment techniques showed that optical techniques will need to be both improved in accuracy for cancer detection and more convenient and cost-effective before they will be accepted by the wider clinical community (St John *et al* 2017). Additionally, new work is being done to achieve pathology-like images through staining resected samples and performing optical imaging; for instance light-sheet microscopy (Glaser *et al* 2017) and fluorescence imaging (Davis *et al* 2013). These show great promise for identifying tumor margins, but in comparison to other optical techniques, have the drawback of requiring tissue staining prior to imaging.

Taking advantage of the autofluorescence properties of breast tissue, earlier studies have shown that fluorescence intensity-based spectroscopy techniques enable detection of breast cancer with good sensitivity and specificity (85% and 96%, respectively) (Keller *et al* 2010). Time-resolved (lifetime) fluorescence spectroscopy techniques can improve these statistics by providing an additional means to analyze tissue autofluorescence by separating tissue fluorophores with overlapping fluorescence intensity parameters but distinct fluorescence lifetimes. Such techniques, however, have only been sparsely explored for diagnosis of breast cancer (Sharma *et al* 2012, Gorpas *et al* 2015). Endogenous fluorophores distinguishable by fluorescence lifetime techniques and relevant to breast cancer detection include adipose tissue, collagen fibers, nicotinamide adenine dinucleotide (NADH) and flavin adenine dinucleotide (FAD).

While the potential diagnostic capabilities of fluorescence lifetime techniques have been demonstrated in pre-clinical studies, many challenges exist for clinical translation including complex instrumentation, time-consuming data analysis, and a lack of ability for clinicians to simply obtain fluorescence lifetime data and quickly display conclusive diagnostic information. Recent advances in fluorescence lifetime imaging (FLIm) instrumentation with a fast and compact scanning fiber-based system (Yankelevich *et al* 2014, Ma *et al* 2015) enable acquisition of FLIm images either during surgery or on excised specimens, in real-time as the FLIm fiber optic is scanned over the tissue via hand scanning or with an automated mechanical stage. The system is housed in a compact cart that can be transported easily to operating or pathology rooms. The goals of this study were to demonstrate: (1) the ability of this compact system to acquire data from breast specimens in scenarios that mimic the intraoperative setting, which would require hand scanning during surgery or a mechanical stage for scanning excised specimens; and (2) the accuracy of a classification algorithm that employs optical parameters derived from FLIm measurements to automatically output diagnostic information about breast specimens as independently validated with histology. Our findings show that this FLIm technique may be a contender for reducing breast cancer re-excision rates due to its ability to accurately and quickly distinguish cancer from normal tissue in a manner that could identify positive margins intraoperatively either during surgery or on resected tissue specimens.

Methods

Breast specimens

Tissue specimens ($N = 20$) from breast cancer patients ($N = 14$ total: $N = 4$ lumpectomies, $N = 10$ mastectomies) were imaged within an hour of resection. Multiple pieces of tissue were imaged from $N = 5$ of the total patients, which is why there are $N = 20$ specimens, but only $N = 14$ patients. All patients provided informed consent. See table 1 for a summary of patient information. The University of California Davis Health System Institutional Review Board approved this study.

Imaging protocol

The tissue was assessed by a pathologist and regions thought to contain tumor were cut into sizes that could be fit in a single tissue processing cassette ($\sim 20\text{ mm} \times 20\text{ mm} \times 4\text{--}5\text{ mm}$), with slight irregularities in the overall shape to assist with later co-registration between histology and FLIm data. Ink was used to mark the edges of the specimen and to assist with co-registration. The samples were placed on an imaging stage and scanned with the FLIm fiber optic either manually by hand or automatically by a mechanical stage to mimic how this system could be used either during surgery or following surgery on excised tissue specimens, respectively. High-

Table 1. Demographic and tumor characteristics of $N = 14$ patients in this study.

Characteristic		Number (%)
Age	<50	3 (21)
	>50	9 (64)
	Unknown	2 (14)
Race	White, not hispanic or latino	12 (86)
	Unknown	2 (14)
Body mass index, kg m ²	Normal (<25)	5 (36)
	Overweight, 25–30	5 (36)
	Obese, ≥30	2 (14)
	Unknown	2 (14)
Menopausal status	Premenopausal	4 (29)
	Postmenopausal	10 (71)
	Unknown	2 (14)
Radiotherapy	No radiotherapy	8 (57)
	Radiotherapy	2 (14)
	Unknown	4 (29)
Hormone therapy	No	7 (50)
	Yes	4 (29)
	Unknown	3 (21)
Type of surgery	Lumpectomy	3 (21)
	Mastectomy	9 (64)
	Unknown	2 (14)
Receptor status	Negative	2 (14)
	Positive	8 (57)
	Unknown	4 (29)
Cancer stage	DCIS	3 (21)
	Invasive	9 (64)
	Unknown	2 (14)

resolution white-light images as well as the video stream of the scanning were also acquired (see **videos 1** and **2** (stacks.iop.org/PMB/63/015003/mmedia)). The fluorescence lifetime values derived from FLIm measurements were augmented with the video stream of the tissue for visualization during imaging and saved for further analysis. The FLIm system and the process to augment the video stream are described below. Following imaging, specimens were placed in formalin and processed routinely for histologic analysis.

Histology

Tissue sections were cut parallel to the imaging plane, thus each histologic section corresponded to one entire field of view of a FLIm dataset. The sections were stained with hematoxylin and eosin (H&E) and scanned with an Aperio Digital Pathology Slide Scanner (Leica Biosystems). The pathologist (M.D.) traced regions of fibrous tissue, normal ducts and lobules, fat, invasive cancer and ductal carcinoma *in situ* (DCIS) using Aperio ImageScope (Leica Biosystems). The FLIm interrogation depth is $\sim 300\ \mu\text{m}$ (Ghosh *et al* 2001, Palmer *et al* 2006) and the depth of a single histologic tissue section was $4\ \mu\text{m}$. To determine how much the breast tissue composition changed within the $300\ \mu\text{m}$ depth, in $N = 2$ cases multiple $4\ \mu\text{m}$ sections were cut within the $300\ \mu\text{m}$ imaged volume. Matlab (The Mathworks, Inc.) software was used for selecting regions of interest (ROIs) in the FLIm images and for image analysis.

Region of Interest Selection

Pathologist tracings from the histology sections were exported from the Aperio software and co-registered with the white light images of the breast tissue, using the shape of the tissue sections and ink as fiducial markers. ROIs were drawn within the tracings, with a $0.5\ \text{mm}$ margin to account for errors in co-registration. See figure 1.

FLIm system

The imaging setup consisted of a prototype point scanning FLIm instrumentation and an aiming beam module (Gorpas *et al* 2016). The aiming beam detection scheme allowed the FLIm images to be reconstructed from the scanning point measurements in real time. The FLIm system is based on a pulse-sampling fluorescence




Figure 1. Representative breast specimen automatically scanned on the mechanical stage. (A) White light image of a breast specimen scanned on a mechanical stage. (B) White light image augmented with FLIm data from spectral channel 2; (C) white light image augmented with classification results when this specimen was left out of the training set. (D) Corresponding H&E histology section. Cancer is outlined in red, adipose in blue and fibrous in green. Scale bar = 4 mm. (E) Zoomed in histology section from black dashed line in (D) with ROIs included in the study shown with the filled-in shapes (red for cancer, blue for adipose, green for fibrous). Scale bar = 0.5 mm. These regions are overlaid with the breast specimen in (F).

lifetime measurement technique and has been described previously (Yankelevich *et al* 2014, Gorpas *et al* 2016). Fluorescence excitation was produced with a micro Q-switched laser frequency tripled to 355 nm with a 2 KHz repetition rate (Teem PhotonicsTM, France). The resulting fluorescence emission from the tissue specimens was sequentially spectrally resolved into four channels: 390/40 nm (channel 1), 466/40 nm (channel 2), 542/50 nm (channel 3), and 629/53 nm (channel 4) (Yankelevich *et al* 2014). Each channel was connected to an optical fiber of varying length that allowed all 4 signals generated from a single laser pulse to arrive sequentially at distinct time points at the detector, a single microchannel plate photomultiplier tube (MCP-PMT, R3809U-50, Hamamatsu, 45 ps FWHM). The signals were then increased by an RF amplifier (AM-1607-3000, 3 GHz bandwidth, Miteq, USA) and temporally resolved (80 ps intervals) by a high sampling frequency digitizer (PXIe-5185, National Instruments, 12.5 GS/s sampling rate). A continuous-wave solid state laser (450 nm, 50 mW, World Star Tech, Canada) coupled into the second channel allowed the aiming beam (power ~ 3 mW) to be projected onto the tissue in the same location as the fluorescence excitation beam. An external camera (Point Grey Chameleon3 1.3 MP Color USB3 Vision with Fujinon HF9HA-1B 2/3" 9 mm lens) recorded the entire specimen, including the aiming beam, during the scanning procedure. The video images were converted to the HSV color space and the blue aiming beam was segmented by thresholding the hue channel, providing co-registration between the FLIm measurements and the video of the tissue. Once the location of the aiming beam is determined, the FLIm data acquired from that location was augmented in real time with the video display of the scanning procedure. Thus as the tissue was scanned, an image of the FLIm data was reconstructed within the video stream of the tissue visualized on the FLIm system computer monitor, creating an augmented view of the tissue overlaid with the FLIm values. (Gorpas *et al* 2016) This can be observed in figures 1 and 2 and videos 1 and 2.

FLIm parameters

Following the acquisition of the fluorescence decay signal, constrained least-squares deconvolution based on the Laguerre expansion method was performed to determine the fluorescence response of the tissue (Liu *et al* 2012). From the deconvolved fluorescence decay, the average lifetimes and intensity ratios were derived. The average lifetime is the average amount of time a fluorophore spends in the excited state. The probability distribution of detected photons is obtained by normalizing the deconvolved fluorescence intensity decay. The average lifetime is then defined as the expected value of this distribution (Lakowicz 1999). Intensity ratios were computed by taking the ratio of the fluorescence intensity at each channel divided by the sum of all four intensity channels.




Figure 2. Representative breast specimen manually scanned by hand. (A) H&E histology section from breast specimen overlaid with pathologist tracings and the ROIs selected for the study (filled-in shapes). Scale bar = 4 mm. (B) The corresponding white light image of the breast specimen augmented with the ROIs identified by the pathologist tracings. (C) White light image augmented with FLIm data from detection channel 2; (D) white light image augmented with the classification results when this specimen was left out of the training set.

Table 2. Results from leave-one-out cross validation with a Gaussian SVM (numbers of pixels per group imbalanced).

	Accuracy ^a	PPV ^a	Sensitivity analysis			Specificity analysis		
			SE ^a	No. TP	FN	SP ^a	No. TN	FP
Cancer	99.0 (99.0–99.1)	98.5 (98.4–98.5)	100.0 (100.0–100.0)	70 555	1	97.4 (97.3–97.5)	40 819	1093
Adipose	99.3 (99.3–99.4)	99.9 (99.9–100.0)	97.2 (97.1–97.3)	25 250	723	100.0 (100.0–100.0)	86 479	16
Fibrous	99.7 (99.6–99.7)	100.0 (100.0–100.0)	97.6 (97.5–97.7)	15 552	387	100.0 (100.0–100.0)	96 527	2

Abbreviations: PPV, positive predictive value; SE, sensitivity; TP, true positive; FN, false negative; SP, specificity; TN, true negatives; FP, false positives.

^a Value expressed as: % (95% CI).

Statistics

Support vector machines (SVM) with a RBF kernel (Chang and Lin 2011) were used to classify FLIm data into three groups based on training from histology: adipose, fibrous and cancerous. The feature vector included average fluorescence lifetime from channels 1–4. Multiclass classification was realized through the ‘one-against-one’ strategy (Hsu and Lin 2002). The cancerous regions included both invasive cancer and DCIS. Sensitivity, specificity, positive predictive value and negative predictive value were calculated with leave one out cross-validation. This involved sequentially leaving data from a single patient out of the training set, then testing the classification accuracy on that single patient for all patients. Since multiple specimens were imaged for $N = 5$ patients, the leave one out cross-validation was performed per patient rather than per specimen. The leave one out cross-validation was performed twice, first with the numbers of pixels per group in the training set imbalanced and next with balanced numbers between groups. The numbers of pixels per group were forced to be balanced by randomly sampling 31 pixels per group from each sample, the size of the smallest group per sample in the dataset (Chawla *et al* 2004). Average fluorescence lifetime values are presented as mean $\pm$ standard deviation. To remove dependence between pixels, the median from each patient from each group was used as the outcome variable and a non-parametric Kruskal–Wallis test was performed to determine statistical significance between groups because the data was not normally distributed, as determined with a Kolmogorov–Smirnov test. Post-hoc Mann–Whitney U -tests were performed to determine the p values for the outcome variables (median values) from each set of groups. Image analysis, classification and statistical analyses were performed using MATLAB (The Mathworks, Inc.). The classification algorithm and results (tables 2 and 3) only included data from a 0.5 mm border within the pathologist tracings of the histology. Data that was scanned by hand was thresholded to remove artifacts that occur at the edges of the specimen (see figure 2). The classification algorithm was validated for the ROIs carefully co-registered with histology, however classification was also performed for all pixels acquired for each specimen (figures 1(C) and 2(D)).

Results

See **videos 1** and **2** for a demonstration of data being recorded and simultaneously displayed on the video feed of samples imaged by hand and by the automated stage, respectively. ROIs were selected from each of the specimens

Table 3. Results from leave-one-out cross validation with a Gaussian SVM (numbers of pixels per group balanced with 31 pixels per group from each sample).

	Accuracy ^a	PPV ^a	Sensitivity analysis		Specificity analysis			
			SE ^a	No. TP	FN	SP ^a	No. TN	FP
Cancer	97.8 (96.9–98.5)	93.1 (91.7–94.4)	100.0 (99.7–100.0)	434	0	96.9 (95.8–97.7)	991	32
Adipose	99.9 (99.6–100.0)	100.0 (100.0–100.0)	99.8 (99.4–100.0)	526	1	100.0 (99.7–100.0)	930	0
Fibrous	97.7 (96.8–98.4)	99.8 (99.3–100.0)	93.5 (92.1–94.7)	464	32	99.9 (99.5–100.0)	960	1

Abbreviations: PPV, positive predictive value; SE, sensitivity; TP, true positive; FN, false negative; SP, specificity; TN, true negatives; FP, false positives.

^a Value expressed as: % (95% CI).

for a total of $N = 14\,688$ pixels associated with fibrous tissue, $N = 67\,465$ associated with cancerous tissue and $N = 24\,311$ associated with adipose tissue. Average spatial resolution was approximately 60 points mm^{-2} .

Average fluorescence lifetime

The mean average lifetime values for each spectral detection channel from within 0.5 mm borders of the pathologist tracings on the FLIm images co-registered with histology (figure 3) were computed. Average fluorescence lifetime from detection channel 1 identifies fibrous regions with the highest values and adipose with lowest values. For spectral channels 2–4 the lifetimes from adipose are highest, fibrous in the middle and cancer the lowest. Figures 1 and 2 demonstrate representative examples. The Kruskal–Wallis test found that the fluorescence lifetimes were significantly different ($p < 0.001$). Additionally, the rank sum test found that the differences in all detection channels between all groups were statistically significantly different ($p < 0.001$) for all groups excluding fibrous compared to cancer in channel 1 ($p = 0.64$).

Classification results

The sensitivity, specificity, positive and negative predictive values for discriminating between adipose, cancerous and fibrous tissue are summarized in table 2. When groups were forced to be balanced by randomly sampling $N = 31$ points per group per sample, the results were slightly different, as summarized in table 3. The SVMs for lifetime values of each set of groups (adipose versus fibrous, adipose versus cancer, fibrous versus cancer) and scatter plots of the fluorescence lifetime data can be seen in figure 4.

Histology co-registration

The pathologist (M.D.) compared histology sections from 3 levels within the 300 μm imaged region and found that the breast tissue did not vary significantly in these $N = 2$ samples to warrant cutting multiple levels from each sample. Thus for the remaining $N = 18$ samples, the first complete section from the paraffin block was used to interpret the results of the entire 300 μm imaged volume.

Discussion

This study demonstrates that spectroscopic features derived from FLIm images are capable of being used to distinguish between adipose, fibrous and cancerous regions in breast specimens from women undergoing lumpectomies and mastectomies. The system is compatible with intraoperative applications. It allows for hand scanning the surgical bed with a fiber optic or automatic scanning of resected *ex vivo* tissue specimens on a mechanical stage. Fluorescence lifetime information is displayed as the scanning is conducted and a classification algorithm was developed to automate distinction between these three tissue types. The classification algorithm worked equally well for data acquired via hand scanning and automated stage scanning. The classification is able to be performed fast enough that it could be implemented in real time as the measurements are acquired, which shows the potential of this technology as an intraoperative tool either during surgery or on resected specimens for tumor margins assessment.

The ability of this FLIm technique to distinguish between breast tissue types is due to the endogenous fluorescence of the fluorophores that comprise those tissues, specifically: fat cells, collagen fibers, NADH and FAD. Adipose tissue is connective tissue predominantly composed of fat cells. From our histologic co-registration, we see that the adipose tissue tends to fluoresce with long lifetimes at the longer wavelengths detected in spectral channels 2–4 of the FLIm apparatus (figure 3), as is consistent with previous studies of adipose tissue fluorescence (Swatland 1987, Datta *et al* 2015). Fibrous tissue is composed of bundles of collagen fibers, and fluoresces with a lifetime longer than cancer, but shorter than adipose tissue, based on our histologic co-registration (figure 3). Importantly, this FLIm technique can clearly distinguish between fibrous and cancer, unlike some other




Figure 3. Average fluorescence lifetime from adipose, fibrous and cancerous breast tissue. This plot includes all data from the ROIs co-registered with histology and included in the classification algorithm. Fibrous: green circles, adipose: blue squares, and cancerous: red diamonds. Fluorescence lifetime (ns) can be seen to vary between the 3 breast tissue types. $P < 0.001$ except for between fibrous and cancer in channel 1.




Figure 4. SVM plots demonstrating discrimination between groups. These plots demonstrate the SVMs that separate the three groups: (A) adipose (blue) and fibrous (green), (B) fibrous and cancer (red), (C) adipose and cancer. The axes represent fluorescence lifetime (ns) in detection channels 1–3.

optical techniques (figures 3 and 4(B)). Cancer cells have altered NADH and FAD metabolism in comparison to normal tissue according to the Warburg theory (Druzhkova *et al* 2016). Free NADH and bound FAD have relatively short lifetimes and fluoresce predominantly in the wavelengths detected by channels 2–4, which may explain the shorter lifetimes of cancerous breast tissue in these channels, though we cannot determine NADH and FAD presence with histologic methods (Skala *et al* 2007). While the trends in fluorescence lifetime detected in spectral channel 1 were also statistically significant, channels 2–4 exhibit the greatest amount of separation between groups.

The margin of healthy tissue around the lumpectomy specimens necessary for negative margins has been controversial. As recently as 2013, standard of care required breast cancer margins to be 1–2 mm in depth, depending on the cancer type and surgeon. However, recent studies demonstrated that a ‘no ink on tumor’ margin for lumpectomy specimens leads to patient outcomes equivalent to those from the previous 1–2 mm guideline (Moran *et al* 2014). The ‘no ink on tumor’ guideline defines negative cancer margins as occurring when there are no tumor cells touching the ink used to mark the entire lumpectomy specimen. The 355 nm FLIm imaging system excitation light penetrates approximately 300 microns into the breast tissue and thus will identify cancer cells right at the surface, in the same region as the ‘no ink on tumor’ guideline suggests. Thus this FLIm system can probe the same region of tissue important for determining margin status based on current clinical guidelines.

The scanning speed of the FLIm system can be varied based on parameters input to the mechanical stage or by the person performing the hand scanning. Two representative speeds can be seen in **videos 1 and 2**. While mechanical stage scanning is slower in this case, the resolution is higher in comparison to the hand scanning, which is faster but with lower resolution. Thus we anticipate hand scanning would be performed to identify positive margins in the operating room or frozen section room immediately following tissue resection and mechanical stage scanning would be reserved for cases where higher resolution is necessary, such as if in the future it is determined that FLIm can be used to study specifics of cancer type as well as positive or negative margins. However, the limiting factor of the slow mechanical stage scanning was the stage itself; with an improved stage, the

FLIm system would be capable of operating at much faster speeds that could allow for the higher resolution scans to be obtained from both the operating room and frozen section room.

We acknowledge that a limitation to this study is the fact that only small regions of tissue within the traced outlines of each tissue type were included in the classification analysis (see figures 1(F) and 2(A)). This study design was used to reduce errors caused by co-registration between the FLIm data and histology that are predominantly caused by: (1) tissue shrinkage and warping during histological processing, (2) the use of a single 4 μm histology section to represent the entire imaged volume and (3) the use of a computer mouse to trace the relevant tissue types, which does not allow for very detailed lines to be drawn with the Aperio software. Our work currently does not take into account the ways that multiple light scattering events may affect fluorescence lifetime, however our excitation-collection geometry and fast temporal resolution reduce the possibility that scattering affects our fluorescence lifetime measurements. In the future, this could be thoroughly studied with a technique such as lifetime tomography (Gao *et al* 2008, Cai *et al* 2016). Additionally, while we hypothesize that NADH and FAD allow us to distinguish cancerous from fibrous and adipose tissue based on known fluorescence lifetime properties of these molecules, we cannot verify this without chemical analysis. Since our specimens are *ex vivo*, this also only allows us to assess the state that the NADH and FAD were in following tissue resection in comparison to the true state they existed in in living tissue. We also acknowledge that these results will need to be validated in a larger cohort, and we aim to move to entire lumpectomy specimens rather than sections of lumpectomy and mastectomy specimens for this work. A larger cohort will also enable the study of whether DCIS and invasive cancer can be distinguished using this technique. In the current study, nevertheless, cancerous tissue, regardless of cancer type, was discriminated from normal breast tissue (both fibrous and adipose).

In conclusion, normal fibrous and adipose tissue was able to be distinguished from cancerous breast tissue with accuracy >97% with a classification algorithm designed using FLIm derived parameters. The FLIm measurements can be acquired within minutes either by hand or automated scanning of a fiber optic, without the need for contrast agents or dyes and without damaging tissue. Additionally, the FLIm signal is generated from the region of breast specimens appropriate for current guidelines for determining tumor margin status. All combined, these results indicate that the current technique has great potential for further application in the field of surgical breast oncology to reduce rates of re-excision by determining tumor margin status intraoperatively either during surgery or on resected tissue specimens.

ORCID iDs

Jennifer E Phipps  <https://orcid.org/0000-0003-0004-2908>

Dimitris Gorpas  <https://orcid.org/0000-0001-9576-8087>

References

- Allen W M, Chin L, Wijesinghe P, Kirk R W, Latham B, Sampson D D, Saunders C M and Kennedy B F 2016 Wide-field optical coherence micro-elastography for intraoperative assessment of human breast cancer margins *Biomed. Opt. Express* **7** 4139–53
- Azu M, Abrahamse P, Katz S J, Jagsi R and Morrow M 2010 What is an adequate margin for breast-conserving surgery? Surgeon attitudes and correlates *Ann. Surg. Oncol.* **17** 558–63
- Brown J Q *et al* 2010 Optical assessment of tumor resection margins in the breast *IEEE J. Sel. Top. Quantum Electron.* **16** 530–44
- Cai C, Cai W, Cheng J, Yang Y and Luo J 2016 Self-guided reconstruction for time-domain fluorescence molecular lifetime tomography *J. Biomed. Opt.* **21** 126012
- Chang C C and Lin C J 2011 LIBSVM: a library for support vector machines *ACM Trans. Intell. Syst. Technol.* **2**
- Chawla N V, Japkowicz N and Kotcz A 2004 Editorial: special issue on learning from imbalanced data sets *SIGKDD Explor. Newsl.* **6** 1–6
- Datta R, Alfonso-Garcia A, Cinco R and Gratton E 2015 Fluorescence lifetime imaging of endogenous biomarker of oxidative stress *Sci. Rep.* **5** 9848
- Davis S C, Gibbs S L, Gunn J R and Pogue B W 2013 Topical dual-stain difference imaging for rapid intra-operative tumor identification in fresh specimens *Opt. Lett.* **38** 5184–7
- de Boer L L, Hendriks B H W, van Duijnhoven F, Peeters-Baas M J T F D V, Van de Vijver K, Loo C E, Jozwiak K, Sterenborg H J C M and Ruers T J M 2016 Using DRS during breast conserving surgery: identifying robust optical parameters and influence of inter-patient variation *Biomed. Opt. Express* **7** 5188–200
- Druzhkova I N, Shirmanova M V, Lukina M M, Dudenkova V V, Mishina N M and Zagaynova E V 2016 The metabolic interaction of cancer cells and fibroblasts—coupling between NAD(P)H and FAD, intracellular pH and hydrogen peroxide *Cell Cycle* **15** 1257–66
- Erickson-Bhatt S J *et al* 2015 Real-time imaging of the resection bed using a handheld probe to reduce incidence of microscopic positive margins in cancer surgery *Cancer Res.* **75** 3706–12
- Fitzmaurice C *et al* 2016 Global, regional, and national cancer incidence, mortality, years of life lost, years lived with disability, and disability-adjusted life-years for 32 cancer groups, 1990 to 2015: a systematic analysis for the global burden of disease study *JAMA Oncol.* **3** 524–48
- Gao F, Zhao H, Zhang L, Tanikawa Y, Marjono A and Yamada Y 2008 A self-normalized, full time-resolved method for fluorescence diffuse optical tomography *Opt. Express* **16** 13104–21
- Ghosh N, Mohanty S K, Majumder S K and Gupta P K 2001 Measurement of optical transport properties of normal and malignant human breast tissue *Appl. Opt.* **40** 176–84
- Glaser A K, Reder N P, Chen Y, McCarty E F, Yin C, Wei L, Wang Y, True L D and Liu J T C 2017 Light-sheet microscopy for slide-free non-destructive pathology of large clinical specimens *Nature Biomed. Eng.* **1** 0084

- Gorpas D, Fatakdawala H, Zhang Y H, Bold R and Marcu L 2015 Fluorescence lifetime spectroscopy for breast cancer margins assessment *Proc. SPIE Optical Biopsy XIII: Toward Real-Time Spectroscopic Imaging and Diagnosis* **9318** 93180B
- Gorpas D, Ma D, Bec J, Yankelevich D and Marcu L 2016 Real-time visualization of tissue surface biochemical features derived from fluorescence lifetime measurements *IEEE Trans. Med. Imaging* (<https://doi.org/10.1109/TMI.2016.2530621>)
- Houssami N, Macaskill P, Marinovich M L and Morrow M 2014 The association of surgical margins and local recurrence in women with early-stage invasive breast cancer treated with breast-conserving therapy: a meta-analysis *Ann. Surg. Oncol.* **21** 717–30
- Hsu C W and Lin C J 2002 A comparison of methods for multiclass support vector machines *IEEE Trnas. Neural Netw.* **13** 415–25
- Keller M D, Majumder S K, Kelley M C, Meszoely I M, Boulos F I, Olivares G M and Mahadevan-Jansen A 2010 Autofluorescence and diffuse reflectance spectroscopy and spectral imaging for breast surgical margin analysis *Laser Surg. Med.* **42** 15–23
- Kennedy S, Caldwell M, Bydlon T, Mulvey C, Mueller J, Wilke L, Barry W, Ramanujam N and Geradts J 2016 Correlation of breast tissue histology and optical signatures to improve margin assessment techniques *J. Biomed. Opt.* **21** 66014
- Kong K, Zaabar F, Rakha E, Ellis I, Koloydenko A and Notingher I 2014 Towards intra-operative diagnosis of tumours during breast conserving surgery by selective-sampling Raman micro-spectroscopy *Phys. Med. Biol.* **59** 6141–52
- Lakowicz J R 1999 *Principles of Fluorescence Spectroscopy* (New York: Springer) (<https://doi.org/10.1007/978-0-387-46312-4>)
- Liu J, Sun Y, Qi J and Marcu L 2012 A novel method for fast and robust estimation of fluorescence decay dynamics using constrained least-squares deconvolution with Laguerre expansion *Phys. Med. Biol.* **57** 843–65
- Ma D, Bec J, Gorpas D, Yankelevich D and Marcu L 2015 Technique for real-time tissue characterization based on scanning multispectral fluorescence lifetime spectroscopy (ms-TRFS) *Biomed. Opt. Express* **6** 987–1002
- Moran M S et al 2014 Society of Surgical Oncology-American Society for Radiation Oncology consensus guideline on margins for breast-conserving surgery with whole-breast irradiation in stages I and II invasive breast cancer *Ann. Surg. Oncol.* **21** 704–16
- Morrow M, Jagsi R, Alderman A K, Griggs J J, Hawley S T, Hamilton A S, Graff J J and Katz S J 2009 Surgeon recommendations and receipt of mastectomy for treatment of breast cancer *JAMA* **302** 1551–6
- Nguyen F T, Zysk A M, Chaney E J, Kotynek J G, Oliphant U J, Bellafiore F J, Rowland K M, Johnson P A and Boppart S A 2009 Intraoperative evaluation of breast tumor margins with optical coherence tomography *Cancer Res.* **69** 8790–6
- Nichols B S et al 2017 Miniature spectral imaging device for wide-field quantitative functional imaging of the morphological landscape of breast tumor margins *J. Biomed. Opt.* **22** 26007
- O'Kelly Priddy C M, Forte V A and Lang J E 2015 The importance of surgical margins in breast cancer *J. Surg. Oncol.* **113** 256–63
- Palmer G M, Zhu C, Breslin T M, Xu F, Gilchrist K W and Ramanujam N 2006 Monte Carlo-based inverse model for calculating tissue optical properties. Part II: application to breast cancer diagnosis *Appl. Opt.* **45** 1072–8
- Sharma V, Shivalingaiah S, Peng Y, Euhus D, Gryczynski Z and Liu H 2012 Auto-fluorescence lifetime and light reflectance spectroscopy for breast cancer diagnosis: potential tools for intraoperative margin detection *Biomed. Opt. Express* **3** 1825–40
- Skala M C, Riching K M, Gendron-Fitzpatrick A, Eickhoff J, Eliceiri K W, White J G and Ramanujam N 2007 *In vivo* multiphoton microscopy of NADH and FAD redox states, fluorescence lifetimes, and cellular morphology in precancerous epithelia *Proc. Natl Acad. Sci. USA* **104** 19494–9
- St John E R, Al-Khudairi R, Ashrafi H, Athanasiou T, Takats Z, Hadjiminas D J, Darzi A and Leff D R 2017 Diagnostic accuracy of intraoperative techniques for margin assessment in breast cancer surgery: a meta-analysis *Ann. Surg.* **265** 300–10
- Swatland H J 1987 Autofluorescence of adipose-tissue measured with fiber optics *Meat Sci.* **19** 277–84
- Wong T T W, Zhang R, Hai P, Zhang C, Pleitez M A, Aft R L, Novack D V and Wang L V 2017 Fast label-free multilayered histology-like imaging of human breast cancer by photoacoustic microscopy *Sci. Adv.* **3** e1602168
- Yankelevich D R, Ma D, Liu J, Sun Y, Sun Y, Bec J, Elson D S and Marcu L 2014 Design and evaluation of a device for fast multispectral time-resolved fluorescence spectroscopy and imaging *Rev. Sci. Instrum.* **85** 034303
- Zysk A M et al 2015 Intraoperative assessment of final margins with a handheld optical imaging probe during breast-conserving surgery may reduce the reoperation rate: results of a multicenter study *Ann. Surg. Oncol.* **22** 3356–62