



**S1E**

Passage number of cell lines: P12

Test method: Biochemical luminescence MycoAlert™ Plus Mycoplasma Detection Kit, Lonza

Line name	Reading A	Reading B	Ratio (B/A)	Result
HMGUi003-A	118	83.2	0.7	Negative for mycoplasma
MRli028-A	98	79.6	0.8	Negative for mycoplasma

Legend:

- Ratio B/A < 1: Negative for Mycoplasma
- Ratio B/A 1- 1.2: Borderline
- Ratio B/A > 1.2: Mycoplasma contamination